


UNDERSTANDING CALIBRATION DOCUMENTS

Application Note: 20210609A

Revision 1.0 (June 2021)

The Calibration Summary Page



The form is a 'CERTIFICATE OF CALIBRATION' from CLIMET INSTRUMENTS COMPANY. It includes fields for instrument model, serial number, calibration date, customer asset number, preparation details, physical condition, calibration status, comments, and signatures/approvals. Callouts [A] point to the asset number field, [B] points to the calibration date and instrument date fields, [C] points to the physical condition section, [D] points to the calibration status section, [E] points to the comments box, and [F] points to the explanatory paragraph at the bottom.

The summary page contains information required by calibration standards, including identification of the instrument, model number, serial number, and the company for whom the instrument was calibrated.

A place is provided [A] for the customer's asset number. This is filled in if we know what the asset number is. (Some companies have more than one asset number on their equipment.)

The date the document was created and the date that the instrument was calibrated is recorded at [B].

The physical condition of the instrument [C] is provided as feedback, since poor packaging can lead to damage in shipment.

The *as found* calibration status [D] is recorded, and the *as left* status: *To specifications*, is checked to provide quick affirmation that the unit met specifications after the calibration was completed.

The *Comments* box [E] is used to note anything relevant to the calibration.

The paragraph below [F] briefly explains parameters calibrated.

This page is signed by the calibration technician and the person that reviews the procedure.

The Calibration Data Page

CLIMET		CERTIFICATE OF CALIBRATION			
CLIMET INSTRUMENTS COMPANY		Calibration Test Data			
C1: _____		Aerosol Particle Counter		S/N: _____	
DATE OF CALIBRATION: _____		TECHNICIAN: _____			
ELECTRONIC MEASUREMENTS					
G	TEST	NOMINAL	TOLERANCE	AS FOUND	PASS
	L.D. DRIVE (VOLTAGE)	Vdc [†]	(reference value)	Vdc	(N/A)
	AIR FLOW	50 LPM	± 2.5 LPM	LPM [‡]	LPM [‡]
H	PEAK NOISE	≤ 200 mV	(reference value)	mV	(N/A)
	* Flow meter reading reflect a corrective offset of $\frac{mV}{LPM}$. LPM = flow measurement in volumetric, corrected for altitude, initial value at the laser diode stage.				
	** Peak noise reading reflect a corrective offset of $\frac{mV}{V}$.				
PERFORMANCE DATA					
NOMINAL PARTICLE SIZE	0.5 μ m	0.8 μ m	1.0 μ m	5.0 μ m	
	EXPECTED AMPLITUDE (from last cal)	mV	mV	V	mV
	TOLERANCE	± 60 mV	± 30 mV	± 165 mV	± 50 mV
	AS FOUND	mV	mV	V	mV
	PASS (Y/N)				
AS LEFT	mV	mV	V	mV	
TEMPERATURE DURING CALIBRATION: _____ °F		HUMIDITY DURING CALIBRATION: _____ %			
ENVIRONMENTAL CONTROLS: Ambient temperature 60°–80°F (18°–26.7°C); humidity has no effect on calibration.					
COLLECTIVE UNCERTAINTY OF MEASUREMENT: ± 2.3% or 0.3 µm and 0.5 µm; ± 3.5% of 5 µm					
The collective uncertainty is based on the contributions of the Pulse Height Analyzer, the Mass Flow Meter and the judgment of the technician in establishing the median of the displayed distribution, as determined by empirical tests and 1 sigma uncertainty calculation.					
ACCURACY RATIO: The collective uncertainty of the measurement standards is less than 25% of the listed tolerances (4:1 measurement ratio).					
CALIBRATION TOLERANCES: The particle size listed are nominal values from the test equipment's record for each test. Tolerance = tolerance times represent a 2% sigma confidence interval for variation from true. If the particle response is below the tolerance for Expected Amplitude the particle will be undercounted, resulting in counts that are less than they would be normally. If the particle response is above the tolerance for Expected Amplitude the particle will be stated larger than it actually is, resulting in counts that are greater than they should be. Accuracy is the ratio of the measured value to the true value. Accuracy is expressed as a percentage of the out-of-tolerance count. Temperature and humidity sensors, if present, are for reference, and are not part of the calibration.					
CALIBRATION PROCEDURE: 92045102: Standard Calibration, T-Series, X-Series, and x5k					
80 UNIT Standard Data, Revised November 4, 2017		Document ID: 123	Certificate of Calibration, Page 3 of 4		

The calibration data page records the ‘as found’ and ‘as left’ measurements made during the calibration. The ‘expected values’ are the ‘as left’ threshold values from the last calibration.

The ‘Tolerances’ provided in the **Performance Data Table** are in mV (millivolts) and reflect a 10% count variance. These tolerances are specified by ISO 21501-4:2018 §6.1.

Preventative Maintenance: Laser Diode drive current [G] is a reference value with no tolerance because the tolerance is in the signal response. It is used by the technician to evaluate the condition of the laser diode. The technician looks for a value that is 20% above the value first reported when the laser diode was installed, but the technician also compares the values recorded during the last two calibrations to spot a high rate of increase that

would predict a laser failure. Laser drive that is within 20% of the original value is still in regulation, and thus does not affect the calibration.

Preventative Maintenance: Peak Noise [H] has no tolerance, but it is recorded for as a reference when evaluating whether the sensor needs cleaning as a preventive maintenance. Increased background noise may make it harder to provide consistent calibrations, and contamination build-up on the mirror (the collection optic) can decrease the amplitude of the signal response, resulting in undercounting.

Below the tables are reports on the limits for environmental factors and the uncertainty of measurement. Since measurements are based on particle sizing, rather than on counts, measurement uncertainty does not translate to counts, and has less effect than count efficiency. Count Efficiency should be used to determine the instrument’s count bias and to evaluate counts reported by the instrument.

The Traceability Page

CLIMET | 1520 WEST COSTA AVENUE, REDLANDS, CA - PHONE (909) 793-2788 • service@climet.com

CERTIFICATE OF CALIBRATION

Standards Traceability

DOCUMENT DATE:

STATEMENT OF TRACEABILITY

This certificate of calibration is issued by the technician listed below. The calibration conforms to Intertek's provisions that meet or exceed the requirements of ISO 21501-4.

Temperature and Relative Humidity are not controlled during calibration because of the wide range of ambient conditions (Temperature: 20°F to 122°F; Humidity: 0-100%, non-condensing).

All test equipment used in the calibration of Climet Instruments' products is calibrated or manufactured recommended intervals by one approved outside source. Climet Instruments' calibration certificates for each piece of test equipment are on file at Climet offices and will be supplied if requested.

Calibration traceability to a National Measurement Institute (NIST). Particle response is established by using reference standards. These spheres are used by method traceable, by lot number, to NIST.

The instruments and reference standards listed below were used to calibrate the instrument certified by this document.

CALIBRATION METHOD

Climet particle counters are calibrated by using one or more sets of reference particles, which serve as standards for comparing and adjusting amplifier response or thresholds of detection to known particle sizes. The particles are introduced to the instrument via a flow system. At a specific concentration, the digital voltmeter is used to make reference measurements. Flow is measured using a Mass Flow Controller (MFC) for the flow rate. An internal or external Pulse Height Analyzer (PHA) measures sensor response to particle challenges. The PHA displays a histogram of the particle distribution, which is used to determine the median of the distribution of particles.

The median of the size which is used to adjust the threshold as needed. To match the threshold or the threshold is adjusted to the median distribution.

The count accuracy is verified during factory prime calibration by count efficiency testing. Count efficiency testing is realized by count comparison using CHC/OMA systems—a primary reference—as a 10/25 count reference.

BENCH # 4

I	INSTRUMENT	ASSET	MODEL	SERIAL NO.	CONTROL NO.	CAL DATE	DUE
	DMA	137	Fluke 87	8924-042	A12388	09 MAY 17	09 JUN 18
	PHA	365	Argus MCA-8000D	01039	01039.10.51.14	24 OCT 16	24 OCT 17
	Mass Flow Meter	199	TD 4240	4040.081.000	0818.002.03.01.17	09 MAR 17	09 MAR 18
	Eh/Temp sensor	111	Vaisala HMP44	U2450023/243038	4718.470.830	30 AUG 17	

PARTICLE STANDARDS

NOMINAL SIZE	ACTUAL SIZE	RSE DEVIATION	LOT NUMBER	EXP. DATE	NOMINAL SIZE	ACTUAL SIZE	RSE DEVIATION	LOT NUMBER	EXP. DATE
0.5 μm	0.503 μm	\pm 0.009 μm			0.6 μm	0.793 μm	\pm 0.009 μm		
0.4 μm	0.400 μm	\pm 0.009 μm			1.0 μm	0.994 μm	\pm 0.015 μm		
0.8 μm	0.808 μm	\pm 0.008 μm			3.0 μm	2.996 μm	\pm 0.004 μm		
0.6 μm	0.600 μm	\pm 0.009 μm			8.0 μm	8.027 μm	\pm 0.047 μm		

[Particle standards used in this calibration are manufactured by Duke Scientific. Only sizes listed with lot number were used in calibration.]

Bench #4 Revision: Revised Sept 21, 2017 Document ID: _____ Certificate of Calibration, Page 3 of 4

The traceability page contains the traceability statement and a brief description of the calibration method. Each technician's bench has its own traceability page.

It contains tables listing the *test equipment* [I] assigned to the bench, and the *test particles* [J] used as part of the calibration.

Test equipment is defined by type of instrument, the asset number and serial number of the instrument, and the model number. The last calibration date and the calibration due date are included, along with the control number of the calibration certificate.

The **particle standards** are identified by nominal size, the actual size, the standard deviation of size, the lot number for traceability, and the expiration date for the particle standards.

The three pages above are included for the ISO 21501-4 compliant calibration.

The Standard calibration is the *size calibration* specified as the calibration method in ISO 21501-4.

This standard also specifies tests to be made following the calibration itself. These tests are reported on a fourth page when the calibration requested is the *ISO calibration*.

The Test Page

CERTIFICATE OF CALIBRATION

Count Efficiency testing

The _____ aerosol particle counter, Serial Number _____ has been checked for 50% and 100% count efficiency by comparison with a Condensation Nucleus Counter (CNC), with particles generated with an Electrostatic Classifier. The results of this count efficiency comparison are listed in the following table:

NOMINAL SIZE	ACTUAL SIZE	SIZE DEVIATION	EFFICIENCY @ SIZE	0.8 µm in 0.5 Ch.
0.5 µm	µm	± nm	(ISO 21501 Spec: 30-70%)	% N/A
0.8 µm	µm	± nm	N/A	(ISO 21501 Spec: 90-110%) %

INSTRUMENT	MODEL	SERIAL NUMBER	ASSET NUMBER	CALIBRATION DATE	DUE DATE
Reference Counter	C-88R	114490	216	08 AUG 2017	08 Aug 2018

Resolution testing

100 LPM 0.5 µm

ISO SPEC: Equal to or less than 15% RESULTS: _____ % PASS

False Count Rate testing

95% UCL count limit: 1 CFM = 32.5 50 LPM = 18.4 75 LPM = 12.3

RESULTS: _____ counts in 10 minutes = _____ counts per cubic meter with 95% UCL PASS

100-0400-001 Rev A, Validated October 4, 2017 Document ID: _____ Calibration Certificate, Page 4 of 4

The test page is supplied with ISO 21501-4 compliant calibrations.

The count efficiency table lists the results of the 50% count efficiency test and the 100% count efficiency test.

The **50% Count Efficiency** test is based on the smallest channel. The threshold for a given channel represents the median of the particle distribution. Under the JIS 9921 standard, this was important, because it assured that 0.1 µm particle counters used in the semiconductor industry were true 0.1 µm instruments, at least within 20 percentage points. ISO 21501-4 sets the acceptable tolerance to $50\% \pm 20\%$. Climet finds this much too loose, and has established our own standard of $\pm 10\%$ for new production particle counters.

The **100% Count Efficiency** test is the important test, because it reports the bias of the instrument, that is, to what degree the instrument undercounts or overcounts. For example, if an instrument has 95% count efficiency, multiplying the counts reported for a given sample by 1.05 will correct for the bias of the instrument, and correcting for the bias of other instruments will provide counts that are more consistent between instruments. The standard allows a variance of $\pm 10\%$ from 100%.

There are no adjustments for fine-tuning count efficiency. The test only verifies that the count efficiency of the instrument meets ISO 21501-4 specifications. While it provides no performance improvement, auditors are increasingly expecting to see evidence of count efficiency testing.

Table [L] reports the results of resolution testing. The limit for resolution is 15%. Resolution is a product of the flow rate of the instrument and the laser diode installed. There is no way to fine tune resolution. Passing resolution assures that particles corresponding to the size of one channel do not overlap the threshold of the next channel and add to the counts in that channel. Because of design choices, it is virtually assured that a Climet particle counter will pass the resolution test.

The False Count Rate (FCR) indicates the potential false counts, with 95% Upper Confidence, that might be reported in a cubic meter sample. ISO 21501-4 does not specify an acceptance criterion. Climet uses ten 1-cubic meter samples during manufacturing to define the FCR. In analyzing the cost benefit of this test in terms of value of the data versus time and expense, Climet has limited this test to 10 minutes for interval calibrations. This keeps costs to the customer down and does not add to delays in returning equipment to the

customer because of the time added to the calibration if longer tests were implemented.

Climet has established a specification of 4 counts in ten minutes in the 0.5 µm channel. Based on a 10-minute sample for a 1 CFM particle counter, the 95% UCL value for 4 counts is 32.5 counts. To put this in perspective, the ISO Class 5 limit for 0.5 µm particles is 3,520 counts, so 32.5 counts represents only 0.92% of the room limit. Even if typical counts were around 1,000, this would only represent a 3% increase in counts.

It is highly unlikely that false counts as high a 4 counts in 10 minutes would violate a room, and if it did, the false count rate should not be the focus of a deviation investigation. A longer test would produce lower 95% UCL counts, but given the limited value of this test, a longer test during the interval calibration would not add value to the calibration.

Testing not included

ISO 21501-4 specifies requirements for the calibration report, but elsewhere it lists requirements for a particle counter that are not tests or are not appropriate at the interval calibration.

- **Maximum particle number concentration ISO 21501-4:2018, §6.5**

The concentration limit is a mathematical calculation made during design. It is a product of the flow rate and the inlet nozzle dimensions. It is not a test. The concentration limit is reported in the User's Manual, not on the calibration report.

- **Sample Time Error ISO 21501-4:2018, §6.7**

In a Climet particle counter, the *sample time* is controlled by the real time clock chip. The chip has an accuracy of at least 1 second in 7 hours. These chips are extremely reliable, and a failure would be obvious and would not require test measurements to detect. Testing the clock circuit is a board level test performed with a frequency counter. This is not a test, but a statement of 'sampling time control system.'

Preventative Maintenance



Climet calibration technicians conduct a number of preventative maintenance tests on each interval calibration:

- **Laser Diode Power
(Ref. Calibration Data Page)**

Climet measures the output current to the laser diode, which is a critical test. This test helps identify Catastrophic Optical Damage, or COD, before it occurs. A COD occurs when high output levels cause a short, melting part of the laser diode edge, and causing a laser diode failure. Please, refer to [Application Note 20210226A](#). A non-authorized 3rd Party Calibration Service Provider does not have access to initial factory values of the instrument, does not have Climet calibration procedures or test points, and cannot perform this test.

- **Peak Noise
(Ref. Calibration Data Page)**

Climet checks the Peak Noise to ensure there's no sensor contamination. High peak noise is a leading indicator of sensor contamination. A non-authorized 3rd Party Calibration Service Provider does not have access to Climet software, and cannot conduct this test.

- **Battery Test**

Climet conducts a capacitive test on battery powered portables to ensure reliable life.

Regardless, we recommend customers implement their own preventative maintenance program and replace batteries every 3-5 years.

- **Memory Battery Replacement Interval**

Depending on the model, Climet will replace the internal coin cell battery at regular intervals.

- **Verification of time and date stamp**

Each calibration includes validation of the time and date stamp for the location where the instrument is received from.

- **Readjustments**

This is the second half of a calibration. Here, the instrument is readjusted annually to ensure high accuracy, high assurance, and high reliability. A non-authorized 3rd Party Calibration Service Provider does not have access to Climet calibration procedures, Climet software, test points, or adjustment values.

- **Validation of Communication Protocols**

Climet validate all Ethernet, RS-232, WiFi, USB, 4-20 mA output signals to ensure proper communication with LIMS equipment.

- **Printer Validation**

Climet technicians test and validate internal printers to identify early or actual failures.

Other Services You May Want To Order

- Exhaust Emissions Test**

This ensure the HEPA filtered exhaust has no leaks, and is doing its job. Climet is the only manufacturers that tests and certifies new production instruments to ISO Class 3 cleanrooms. We recommend you re-test the HEPA filter after 5 years.

- High Pressure Diffuser Cleaning
Maintenance and Testing**

The high pressure diffuser is commonly used in conjunction with particle counters and microbial air samplers to measure inert and viable contamination in high pressure gases. It has no moving parts and does not need a calibration. However, we do recommend a routine cleaning every two years. Please, refer to the User Manual if you wish to clean these instruments yourself.



Copyright © 2021
All Rights Reserved

Climet Instruments Company, a Division of Venturedyne Ltd.

www.climet.com

This publication may contain other company names, and brand names, which may be registered trademarks of their respective holders. Best efforts have been used in this document to present accurate information, but inadvertently may contain typographical errors, omissions, or other inaccuracies. The author disclaims liability from any unforeseen errors and omissions; and we reserve the right to correct any errors, inaccuracies or omissions and to change or update information at any time, without prior notice. We do not generally undertake to update, modify or clarify information on this document, except as required by law. Users may recommend additions, modifications, or deletions by contacting sales@climet.com and referring to this document and revision number.